

ExtraMAG 3 DNA/RNA Isolation Kit


User manual



#EM3.3



LabPack LLC, Address: Karpovka River Embankment, 5, letter I, pom. 3-N, Saint-Petersburg 197376, Russia, phone: +7(812) 490-75-93, e-mail: info@labpack.ru



MagnaLab LLC, Address: Ludwika Rydygiera Str. 8/4, Warsaw 01-793, Poland, phone: +48 531 261 817, e-mail: info@magnalab.pl



Table of Contents

1. INTENDED USE	2
2. PRINCIPLES OF THE PROCEDURE	2
3. KIT DESCRIPTION	2
4. PERFORMANCE CHARACTERISTICS	3
5. CLINICAL MATERIAL RECOMENDATION	3
6. MATERIALS REQUIRED BUT NOT SUPPLIED*	4
7. INSTRUCTIONS FOR USE	4
8. POSSIBLE PROBLEMS AND THEIR SOLUTION	5
9. REAGENT HANDLING AND STORAGE	6
10. WARNINGS AND SPECIAL PRECAUTIONS REGARDING THE SAFE DISPOSAL OF REAGENT KIT	6
11. LIMITATIONS	7
12. MANUFACTURER'S WARRANTIES	7
Appendix A. Symbols to be used with medical device labels	8
Appendix B. Harmonised standarts	9

Abbreviations

RT-PCR	Reverse transcription polymerase chain reaction
Real-time PCR-RT	Real-time polymerase chain reaction with fluorescence detection with reverse transcription
NA	Nucleic acid
PC	Positive control sample
NC	Negative control sample

1. INTENDED USE

ExtraMAG 3 is intended for the isolation of DNA or RNA from clinical material (oropharyngeal swab, nasopharyngeal swab, sputum) by automatic method based on the reversible binding of nucleic acid molecules on the surface of magnetic sorbent particles in the presence of chaotropic salts, for subsequent detection and identification of NA by polymerase chain reaction (PCR) or reverse transcription polymerase chain reaction (RT-PCR), in particular for diagnostics of SARS-CoV-2 infection.

Demographic and population aspects of the application. It is recommended to use ExtraMAG 3 for the preparation of clinical samples obtained from male and female patients without age restrictions.

The intended purpose of ExtraMAG 3 is to provide a preanalytical stage for the analysis of biological samples performed during the clinical laboratory diagnostics of infectious pathology using PCR or RT-PCR. The kit is used to obtain NA samples free of inhibitors of reverse transcription and PCR, thus ensuring high analytical capabilities for subsequent analysis.

The material for the NA extraction procedure is clinical material (oropharyngeal swab, nasopharyngeal swab, sputum).

Application. Professional use in clinical laboratory diagnostics and research practice. For *in vitro* diagnostics only.

Users' qualification requirements. A specialist with a higher or secondary specialized medical education, trained in licensed laboratory diagnostics courses.

2. PRINCIPLES OF THE PROCEDURE

The working principle of ExtraMAG 3 is based on the reversible binding of nucleic acid molecules on the surface of magnetic sorbent particles. The sample is treated with a lysis solution in the presence of magnetic sorbent particles. As a result, cell membranes, viral membranes and other biopolymer complexes are destroyed, and nucleic acids are released. Dissolved NAs link with the sorbent particles, while the other components of the lysed biological material remain in solution and are removed during magnetic precipitation of sorbent on a magnetic stand followed by washing of the sorbent. When an elution buffer is added to the magnetic sorbent, NAs are eluted from the surface of the sorbent into solution, which is then separated from the sorbent particles by magnetic force. As a result of this procedure, a highly purified NA sample free from amplification reaction inhibitors is obtained, which ensures high analytical sensitivity of the PCR or RT-PCR study.

3. KIT DESCRIPTION

ExtraMAG 3 is for automatic use (for example, robotic nucleic acid sample preparation stations Freedom EVO® series, TECAN Schweiz AG or similar).

The reagents that are provided in this kit type are sufficient for 96 reactions. ExtraMAG 3 is not sterile.

Table 3. ExtraMAG 3 description

Kit Component	Description	Amount
Plate with Lysis Buffer and Magnetic sorbent	96-well plate, each well contains 0.4 ml of suspension (black or brown particles in the clear liquid); after the plate stirring within 5–10 seconds – opaque uniform suspension of black or brown color	1 piece
Wash Buffer 1	Colorless clear liquid	40 ml x 2 bottles
Wash Buffer 2	Colorless clear liquid	40 ml x 2 bottles

Elution Buffer	Colorless clear liquid	12 ml x 1 bottle
User manual	-	1 piece
Certificate of quality	-	1 piece

4. PERFORMANCE CHARACTERISTICS

ExtraMAG 3 provides isolation from the clinical material of NAs with a A260/280 purity ratio of at least 1.7.

To control the quality of biomaterial sampling and NA extraction, it is recommended to use the quality controls of biomaterial sampling, which are part of the reagent kits for PCR or RT-PCR as a separate component or use reagent kits for real-time PCR or PCR-RT, which include a fluorescent probe for the detection of endogenous human genomic NA in the sample.

The influence of potentially **interfering substances** on the operation of the reagent kit was tested in relation to potentially interfering substances that will occur during the procedure for collecting biological material:

1. Hemoglobin 10% v/v
2. Mucin 5% v/v
3. "Ibuprofen" 0.04 mg/ml
4. "Ambrobene" 0.003 mg/ml
5. "Bromhexine" 0.016 mg/ml
6. "Kaletra" 0.02 mg/ml
7. "Interferon" 0.2 U/ml
8. "Teraflu" 0.071 mg/ml

These potentially interfering substances encountered in the procedure for isolating NA from clinical material, evaluated at concentrations that would be expected to occur during normal use of the reagent kit, do not interfere with the assay result.

5. CLINICAL MATERIAL RECOMENDATION

Clinical samples, including oropharyngeal swab, nasopharyngeal swab or sputum are to be used as clinical material for the extraction of nucleic acids.

According to WHO guidelines «Diagnostic testing for SARS-CoV-2» the optimal specimen depends on clinical presentation and time since symptom onset. Respiratory specimens:

- Upper respiratory specimens are adequate for testing early-stage infections, especially in asymptomatic or mild cases. Testing combined nasopharyngeal and oropharyngeal swabs from one individual has been shown to increase sensitivity for detection of respiratory viruses and improve the reliability of the result. Two individual swabs can be combined in one collection tube or a combined nasopharyngeal and oropharyngeal swab can be taken. A few studies have found that individual nasopharyngeal swabs yield a more reliable result than oropharyngeal swabs.

- Lower respiratory specimens are advised if collected later in the course of the COVID-19 disease or in patients with a negative URT sampling and there is a strong clinical suspicion of COVID-19. LRT specimens can consist of sputum, if spontaneously produced (induced sputum is not recommended as this poses an increased risk of aerosol transmission). Before implementing other respiratory or oral fluid sampling methods, the sampling method should first pass validation in the laboratory for the intended patient groups.

Specimens for virus detection should reach the laboratory as soon as possible after collection. Correct handling of specimens during transportation and in the laboratory is essential.

Specimen type	Collection materials	Recommended temperature for storage and/or shipment to the laboratory and until testing (from date of specimen collection) #
Nasopharyngeal and oropharyngeal swab	Dacron or polyester flocked swabs with VTM *	2-8 °C if ≤12 days* -70 °C (dry ice) if > 12 days
Sputum	Sterile container	2-8 °C if ≤ 2 days -70 °C (dry ice) if > 2 days

Avoid repeated freezing and thawing of specimens. If no access to $-70\text{ }^{\circ}\text{C}$ consider storing at $-20\text{ }^{\circ}\text{C}$.
 * For transport of specimens for viral detection, use preferentially viral transport medium (VTM) containing antifungal and antibiotic supplements. If VTM is not available, other solutions may be used after validation. Such solution may include phosphate buffered saline (PBS), 0.9% sterile saline, minimum essential medium (with storage at $+4\text{C}$ up to 7 to 14 days). In case other viruses such as influenza should also be tested, do not store samples for more than 5 days at 4-8 degrees but $-70\text{ }^{\circ}\text{C}$ or dry ice.

6. MATERIALS REQUIRED BUT NOT SUPPLIED*

- Laminar box II or III class of biological safety (for example, "BAVp-01-"Laminar-S"-1,2", "Laminar systems", Russia);
- Mini-Centrifuge/Vortex (for example, FV-2400 Micro-Spin, SIA Biosan, Latvia);
- Bench-top centrifuge for of the Eppendorf type microtubes (1.5–2 ml) up to 10000 g (for example, MiniSpin[®], Eppendorf, Germany);
- Robotic nucleic acid sample preparation stations Freedom EVO[®] series from TECAN Schweiz AG, Switzerland complete with consumables;
- Refrigerator with chambers that maintain a temperature of $+2$ to $+8\text{ }^{\circ}\text{C}$ (for storing the "ExtraMAG 3" reagent kit);
- Refrigerator with a chamber that maintains a temperature of $+2$ to $+8\text{ }^{\circ}\text{C}$ (for storing the NA samples). Storage of the NA samples in the same refrigerator with the components of the NA isolation kit is not allowed;
- Eppendorf type microtubes (1.5–2 ml), with Safe-Lock (for example, Axygen, USA);
- Microtube rack (for 1.5–2 ml) (for example, Axygen, USA);
- 1-channel mechanical pipettes with a variable dosing volume of 2–20 μl , 5–50 μl , 20–200 μl , 100–1000 μl certified by the average dose value and the repeatability of pipetting results (error no more than 3%) (for example, Sartorius Biohit, Finland);
- Disposable tips with a filter for semi-automatic pipettes marked "RNAase-free, DNAase-free" in volumes of 1–20 μl , 5–50 μl , 20–200 μl , 100–1000 μl (for example, Axygen, USA);
- Personal protective equipment (PPE) (disposable dressing gown, hat, mask, gloves);
- Container with a lid for a disinfectant solution.

* In the absence of the above equipment and materials, it is allowed to use other means with characteristics no worse than those given.

7. INSTRUCTIONS FOR USE

7.1. Preparation of the components of ExtraMAG 3 for analysis.

Table 2. Preparation of the components for analysis

Kit component	Preparation of the components
Plate with Lysis Buffer and Magnetic Sorbent	Ready to use
Wash Buffer 1	Ready to use
Wash Buffer 2	Ready to use
Elution Buffer	Ready to use

ATTENTION! Before work, it is necessary to read the instruction manual for the Tecan robotic sample preparation station of the Freedom EVO[®] series.

7.2.3.1. Program the device in accordance with the instruction manual for the Tecan Freedom EVO[®] Series Robotic NK Sample Preparation Station (script is provided by the manufacturer upon request).

7.2.3.2. Dispense 100 μl of clinical samples into the wells, 100 μl of PC* into a separate well for the positive control, and 100 μl of NC* into a separate well for the negative control, according to the procedure.

* - included in kits for PCR or RT-PCR.

7.2.3.3. Start the program on the Tecan Freedom EVO device or a similar device (see the operating

instructions for the device) following the device prompts to install the plates in the automatic nucleic acid purification station:

- Plate with Lysis Buffer and Magnetic Sorbent and added samples;
- Bottle with Wash Buffer 1;
- Bottle with Wash Buffer 2;
- Bottle with Elution Buffer.

After the end of the extraction program, remove the plates from the device.

NA samples can be stored for 30 min at temperatures from + 2 to + 8 °C or for 1 week at a temperature not higher than minus16°C.

8. POSSIBLE PROBLEMS AND THEIR SOLUTION

8.1. Absence of a positive reaction with a known positive sample during PCR or RT-PCR.

№	Possible reason	Solution
1	Incomplete cell lysis due to the presence of a crystalline precipitate in the lysis buffer.	Before each isolation procedure, it is necessary to thoroughly mix the Lysis Buffer; if crystals appear, warm the vial at +65 °C until they are completely dissolved.
2	Incorrect sample preparation due to non-compliance with recommendations for the procedure for obtaining clinical samples (including violation of storage and transportation conditions).	Repeat the selection on newly selected samples.
3	Insufficient amount of biological material in the sample.	Repeat the selection on newly selected samples.

8.2. The presence of a positive reaction with a known negative sample during PCR or RT-PCR.

№	Possible reason	Solution
1	Contamination at the stage of NA extraction	Decontaminate, use filter tips, chemical and ultraviolet disinfection of all work surfaces, use separate sets of dispensers, equipment, gowns and gloves for each area, conduct internal and external quality control of studies.

9. REAGENT HANDLING AND STORAGE

9.1. Storage conditions

Store ExtraMAG 3 at temperatures from + 2 to +8 °C. Kits stored in violation of the regulated regime are not subject to use.

9.2. Transportation conditions

ExtraMAG 3 can be transported by covered transport (road, rail or air) in thermal containers containing refrigeration elements at temperatures from +2 to +8 °C. Kits transported in violation of the temperature regime cannot be used.

9.3. Product shelf life

ExtraMAG 3 shelf life is 12 months.

9.4. Storage conditions and shelf life of opened product components.

№	Component	Storage conditions and shelf life
1	Wash Buffer 1	1 month at a temperature from +2 to +8 °C
2	Wash Buffer 2	1 month at a temperature from +2 to +8 °C
3	Elution Buffer	1 month at a temperature from +2 to +8 °C
4	Plate with Lysis Buffer and Magnetic Sorbent	Not stored

10. WARNINGS AND SPECIAL PRECAUTIONS REGARDING THE SAFE DISPOSAL OF REAGENT KIT

10.1. Only personnel trained in molecular diagnostics is allowed to work with the kit of reagent. The rules for working in a clinical diagnostic laboratory have to be observed.

10.2. To prevent contamination of newly investigated samples, reagents and consumables with amplification products (amplicons), nucleic acid preparations or biomaterials, and, as a consequence, the appearance of false positive results, the laboratory process should be unidirectional. Separate rooms (zones) are used for different stages of the analysis. Work should start in the isolation zone, continue in the amplification and detection zones. Do not return samples, equipment and reagents to the area where the previous step of the process was carried out.

10.3. ExtraMAG 3 is intended for single use when the specified number of samples is to be isolated.

10.4. It is allowed to use the kit for the intended purpose only, in accordance with these instructions and within the indicated expiration date.

10.5. Do not use the reagent kit if the inner packaging is broken or the appearance of the reagent does not correspond to the description.

10.6. Do not use the reagent kit if the conditions of transportation and storage have not been observed according to the instructions.

10.7. In the process of work, it is imperative to use personal protective equipment: disposable gloves, laboratory coats. Wash hands thoroughly after finishing work.

10.8. When working with the kit, avoid contact of the kit components with the skin, eyes and mucous membranes. In case of contact, immediately rinse the affected area with water and seek medical attention.

10.9. The components of the kit "Wash Buffer 1", "Wash Buffer 2" isopropyl alcohol and acetone, which are classified as flammable liquids. Electrical equipment and lighting when working with isopropyl alcohol and acetone must be explosion-proof.

10.10. The component of the kit "Plate with Lysis Buffer and Magnetic Sorbent" contains guanidine thiocyanate, which can be absorbed through the skin and is a sensitizing agent. In case of contact with skin or eyes, immediately rinse these areas of the body with water.

10.11. Disposal of the reagent kit. Unused reagents, expired reagents, as well as used reagents, packaging, biological material, including materials, tools and items contaminated with biological material, should be disposed in accordance with the requirements of Directive 2008/98/EC.

The choice of the method of disinfection / neutralization is determined by the capabilities of the organization carrying out medical activities and is carried out when developing a scheme for handling medical waste. In the absence of a disinfection / neutralization site or a centralized medical waste disposal system adopted in the administrative territory in an organization carrying out medical activities; the kits are disinfected by the personnel of this organization in the places of their formation by chemical / physical methods.

The manufacturer, suppliers, sellers, importers can destroy kits that have lost their consumer properties or have expired, in accordance with Directive 2008/98/EC as class H9 medical waste using methods agreed with the territorial authorities responsible for sanitary and epidemiological well-being of the population. The personnel carrying out the disposal of the kits have to comply with the safety rules for carrying out one or another method of destruction.

11. LIMITATIONS

Contamination at the stage of NA isolation is a possible reason for obtaining a false positive result during the subsequent procedure of detecting and identifying NA by PCR or RT-PCR.

12. MANUFACTURER'S WARRANTIES

12.1. ExtraMAG 3 is manufactured by LabPack LLC. Address: Karpovka River Embarkment, 5, letter I, room 3-N, 197376 Saint-Petersburg, Russia.

Phone: +7 (812) 490 75 93, e-mail: info@labpack.ru.

12.2. The manufacturer guarantees the compliance ExtraMAG 3 with the requirements of the technical documentation if the conditions of transportation, storage and use established by this instruction for use are observed.


















12.3. Instruction manual is provided by the manufacturer on paper and in the form of an electronic document via online publishing. The latest revision of the instruction manual: 31/08/2021.

12.4. Complaints about the quality of ExtraMAG 3 should be sent to an authorized representative of the manufacturer in the European Community:

MagnaLab LLC, Address: Ludwika Rydygiera Str. 8/4, Warsaw 01-793, Poland.

Phone: +48 531 261 817, e-mail: info@magnalab.pl

Appendix A. Symbols to be used with medical device labels

Symbol	Title of symbol	Symbol	Title of symbol
EN ISO 15223-1:2016			
	Use-by date		Do not re-use
	Date of manufacture		Consult instructions for use
	Batch code		Caution
	Catalogue number		<i>In vitro</i> diagnostic medical device
	Keep away from sunlight		Do not use if package is damaged
	Keep dry		Contains sufficient for <n> tests
	Temperature limit		Manufacturer
	Authorized representative in the European Community		
Regulation (EC) No 1272/2008			
	H225: Highly flammable liquid and vapour		
	H315: Causes skin irritation		

Appendix B. Harmonised standards

- | | | |
|----|----------------------|--|
| 1 | EN ISO 13485:2016 | Medical devices – Quality management systems – Requirements for regulatory purposes |
| 2 | EN 13612:2002 | Performance evaluation of in vitro diagnostic medical devices |
| 3 | EN 13641:2002 | Elimination or reduction of risk of infection related to in vitro diagnostic reagents |
| 4 | EN 13975:2003 | Sampling procedures used for acceptance testing of in vitro diagnostic medical devices – Statistical aspects |
| 5 | EN ISO 14971:2012 | Medical devices – Application of risk management to medical devices |
| 6 | EN ISO 15223-1:2016 | Medical devices – Symbols to be used with medical device labels, labelling and information to be supplied – Part 1: General requirements |
| 7 | EN ISO 18113-1:2011 | In vitro diagnostic medical devices – Information supplied by the manufacturer (labelling) – Part 1: Terms, definitions and general requirements |
| 8 | EN ISO 18113-2:2011 | In vitro diagnostic medical devices – Information supplied by the manufacturer (labelling) – Part 2: In vitro diagnostic reagents for professional use |
| 9 | EN ISO 23640:2015 | In vitro diagnostic medical devices – Evaluation of stability of in vitro diagnostic reagents |
| 10 | EN 62366-1-2015 | Medical devices – Part 1: Application of usability engineering to medical devices |
| 11 | Directive 98/79/EC | Directive 98/79/EC of the European Parliament and of the Council on in vitro diagnostic medical devices |
| 12 | Directive 2008/98/EC | Directive 2008/98/EC of the European Parliament and of the Council of 19 November 2008 on waste and repealing certain Directives |

The above standards were valid at the time the instructions for use were approved. In the future, when using the document, it is advisable to check the validity of the reference normative documents at the current moment. If the referenced document is replaced or changed, then when applying this document, the replaced (modified) document should be used.